

In vitro Diagnostic medical device

Asan qPCR Test® Dengue Virus

1. LOT number and Expiry date

Refer to external labelling

2. Intended use

The Asan qPCR Test® Dengue Virus is an in vitro diagnostic medical device designed to qualitatively detect the 3'UTR gene of Dengue Virus types 1-4 in plasma (EDTA) and serum samples from individuals suspected of Dengue Virus infection. It utilizes Real-time Reverse Transcription Polymerase Chain Reaction (Real-time RT-PCR) to assist in the diagnosis of Dengue Virus infection.

3. How to use

(1) Product description

This product is a One-step RT-PCR product that uses Real-time reverse transcription polymerase chain reaction (Real-time RT-PCR). One-Step RT-PCR uses materials for reverse transcription reaction and PCR reaction mixed in one tube. In this process, cDNA is synthesized from RNA extracted from the specimen in the same tube, and PCR for the target gene is performed from the synthesized cDNA.

The real-time PCR method uses a primer and probe, wherein the amplification product amplified by the primer binds to the probe, the polymerase decomposes the probe, and the fluorescence signal of the generated reporter dye is measured in real time. In this process, the fluorescence signal increases according to the amount of amplification product. The Human RNase P gene, which is an internal positive control (IC), can check the RNA extraction process, whether PCR reaction inhibitors are mixed in the extracted RNA, and whether the sample was properly collected and extracted.

This product is capable of detecting Dengue virus serotypes 1~4 from RNA extracted from human plasma and serum samples.

(2) Kit Components

The Asan qPCR Test® Dengue Virus kit contains sufficient reagents and quality control samples. The kit contains the following:

	Components	Volume	Quantity
1	Onestep qPCR Master Mix	1.1 mL	1 tube
2	Primer/Probe Mix	400 µL	1 tube
3	Positive Control	250 µL	1 tube
4	Negative Control	250 µL	1 tube

※ Material and devices required but not provided

CFX 96™ Real-Time PCR Detection system(Bio-Rad)

- Nucleic acid extraction system or kit
- Benchtop centrifuge with a rotor for 1.5/2.0 mL reaction tubes
- Centrifuge with a rotor for microtiter plates
- Vortex mixer
- Appropriate reaction plate or tube
- Pipettes
- Pipette tips with filters
- Disposable latex gloves

(3) Specimen preparation and storage

- 1) The specimens that can be used are human plasma(EDTA) and serum samples
 - 2) Specimens stored in transport medium are stable for up to 72 hours in refrigerated storage (2~8°C), and must be stored frozen under -70°C for long-term storage.
 - 3) Test sensitivity may decrease when samples are stored for a long period of time or are repeatedly frozen and thawed.
- ※ All clinical specimens are handled as potentially infectious.

4. Procedure

(1) Nucleic acid extraction

It is recommended to use the QIAamp DSP RNA Viral Mini Kit as a nucleic acid extraction reagent, and follow the manufacturer's instructions.

(2) How to test

[Preparation of reagent]

- 1) Open the box, and thaw each components
- 2) Lightly vortex each completely melted component and spin-down.

[Mix the components]

- 1) After considering the number of samples N (number of specimens + Positive Control 1EA + Negative Control 1EA + Extra 1EA) to be tested, prepare 15 µl x N of reaction solution using Onestep qPCR Master Mix and Primer/Probe Mix as shown in the table below.

	Components	1 Test Volume	N Test Volume
1	Onestep qPCR Master Mix	11 ul	11 ul x N
2	Primer/Probe Mix	4 ul	4 ul x N
3	Viral RNA Sample	5 ul	5 ul x N
	Total	20 ul	20 ul x N

- 2) Vortex the prepared reaction solution lightly, being careful not to create bubbles, and then spin down.
- 3) Dispense 15 µl of the prepared reaction solution into the PCR tube.
- 4) Dispense 5 µl of the extracted Viral RNA Sample, Positive Control (PC), and Negative Control (NC) into each PCR Tube. (To minimize contamination, work in the following order: NC > Sample > PC.

[Amplification of Real-time PCR]

- 1) After dispensing, spin down the PCR tube and place it into the PCR well.
- 2) Operate the equipment following the user manual instructions.
- 3) Refer to the RT-PCR equipment settings in the following table.

<Setting of CFX96™ Dx System fluorescent>

Target	Dengue Virus	IC
Fluorophore	FAM	Cy5

<Setting of Real-time PCR condition>

Steps	Temperature	Time	Cycle
1 cDNA synthesis	50°C	5 min	1
2 RTase Inactivation	95°C	3 min	1
3 Pre-Denaturation	95°C	1 sec	5
4 Pre-Annealing/ Extension	60°C	1 sec	
5 Denaturation	95°C	1 sec	40
6 Annealing / Extension	60°C	1 sec	

5. Result analysis

After the test is completed, open the corresponding experiment data in the analysis program, perform the result analysis, check the Cut-off (Ct value) for the target gene, and set the threshold to 500.

Target	Threshold	Ct Value	Result
Dengue Virus	500	Ct ≤ 36	Dengue Virus positive
IC	500	Ct ≤ 34	Internal control positive

Case	Dengue Virus (3'UTR)	IC	Interpretation
1	+	+/-	Dengue Virus positive
2	-	+	Dengue Virus negative
3	-	-	Invalid/Re-test

- Even if the internal control material (IC) shows a negative result, it may still be judged positive. If each detected gene is strongly positive, the IC result value may be impaired due to the high concentration of the detected gene. Therefore, a positive judgment is made even if the IC is negative.

- Invalid/Re-test: If the test is invalid, a retest is performed.

6. Quality control and validity of results

A minimum of one Negative Control and one Positive Control must be included in each run. All control wells must pass for the Real-time RT-PCR plate to be considered valid.

Target	Dengue Virus (FAM)	IC (Cy5)
PC	Ct ≤ 36	Ct ≤ 34
NC	N/A	N/A

7. Performance

(1) Analytical sensitivity – Limit of detection (LoD)

To determine the LOD, Dengue Virus serotypes 1-4 standards were serially diluted. Each sample was tested 60 times in total, using two lots, with 2 runs per day over 3 days, and 5 tests per run. The lowest concentration at which the virus was detected in more than 95% of the tests was identified.

Detection gene	Limit of detection (LoD)
	Detection concentration (Copies/μL)
DENV-1	10 Copies/μL
DENV-2	10 Copies/μL
DENV-3	8 Copies/μL
DENV-4	8 Copies/μL

(2) Precision - Repeatability

Three samples of high concentration (100xLoD), medium concentration (10xLoD), and low concentration (2xLoD) were tested as a set by the same experimenter, twice daily, three times per test, over five days using one lot. The test confirmed that the Ct values detected within each test, between tests, and across test dates had a coefficient of variation (CV) of less than 5%, indicating effective repeatability.

(3) Precision – Reproducibility

Three samples of high concentration (100xLoD), medium concentration (10xLoD), and low concentration (2xLoD) were used as a set to assess reproducibility across locations, testers, and lots. Two testers across two laboratories tested two different lots twice daily for five days, with each test repeated three times. The results consistently showed 100% agreement between negative and positive outcomes, and the coefficient of variation (CV) for detected Ct values was less than 5%. This confirmed effective reproducibility.

(4) Analytical specificity - cross-reactivity

Samples containing 29 different microorganisms and viruses that could potentially be present were tested to evaluate cross-reactivity. Each sample was tested three times on a single instrument using a single lot of the product. The results showed no cross-reactivity at the analyzed concentrations.

No	Pathogen name	Concentration	Detection or not
1	Zika virus	6.5 ng/ul	N.D.
2	Chikungunya virus	11.4 ng/ul	N.D.
3	West Nile Virus	10 ^{3.8} TCID ₅₀ /mL	N.D.
4	Yersinia enterocolitica	25 ng/ml	N.D.
5	Japanese Encephalitis virus	13.9 ng/ul	N.D.
6	Human cytomegalovirus	5x10 ⁵ pfu/ml	N.D.
7	Bacillus cereus	5 ng/ul	N.D.
8	Clostridium difficile	1x10 ⁵ CFU/ml	N.D.
9	Enterococcus faecalis	5.12 ng/ml	N.D.
10	Escherichia coli	1x10 ⁵ CFU/ml	N.D.
11	Haemophilus parainfluenzae	1x10 ⁵ CFU/ml	N.D.
12	Hemophilus influenzae	1x10 ⁵ CFU/ml	N.D.
13	Enterovirus B111	0.7x10 ⁴ TCID ₅₀ /ml	N.D.
14	Enterovirus 68	4.75x10 ⁵ TCID ₅₀ /ml	N.D.
15	Influenza A H1N1	4.17x10 ⁵ TCID ₅₀ /ml	N.D.
16	Influenza A H3N2	16.4 ng/ul	N.D.
17	Influenza B	102.7 ng/ul	N.D.
18	Japanese Encephalitis virus	10 ³ pfu/ml	N.D.
19	Klebsiella pneumoniae	1x10 ⁵ CFU/ml	N.D.
20	Listeria monocytogenes	1x10 ⁵ CFU/ml	N.D.
21	Mycobacterium fortuitum	8.053x10 ⁷ copies/ul	N.D.
22	Mycobacterium goodii	1x10 ⁵ CFU/ml	N.D.
23	Proreus vulgaris	1x10 ⁶ CFU/ml	N.D.
24	Proteus mirabilis	1x10 ⁵ CFU/ml	N.D.
25	Rotavirus A	1x10 ⁵ copies/ul	N.D.
26	Salmonella Typhimurium	1x10 ⁵ CFU/ml	N.D.
27	Salmonella Typhi	1x10 ⁵ CFU/ml	N.D.
28	Staphylococcus aureus	1x10 ⁵ CFU/ml	N.D.
29	Streptococcus pneumoniae	2.318x10 ⁷ copies/ul	N.D.

N.D: Not detected

(5) Analytical specificity – interference response

Samples were spiked with seven types of endogenous or exogenous substances and tested three times per sample. It was confirmed that there was no interference from any of the seven substances at the tested concentrations.

No	Interference material	Concentration	Interference or not
1	EDTA	292 umol/L	N.A

2	Heparin	3000 U/L	N.A
3	Sodium citrate	12.9 pmol/L	N.A
4	Bilirubin, Conjugated	0.05 mg/mL	N.A
5	Albumin	0.24 g/mL	N.A
6	Haemoglobin	2 mg/mL	N.A
7	Biotin	300 ug/mL	N.A

N.A: Not available

8. Storage method and period of use

(1) Store the kit in its original sealed packaging at -20°C. or below, away from direct sunlight. Kit contents are stable until the expiration date printed on the outer box.

(2) The number of freeze/thaw cycles is limited to four

9. Precautions

[General precautions]

(1) This product is for in vitro diagnostic use only.

(2) This test should only be conducted by trained professionals.

(3) Before starting the assay, thoroughly read the user manual and follow the instructions for use.

(4) All specimens collected from humans should be considered potentially infectious. Follow Good Laboratory Practice (GLP) when handling samples to prevent contamination.

(5) Avoid contact of this product with skin, eyes, or mucous membranes. If contact occurs, rinse immediately with plenty of water.

(6) Improper control of sample cross-contamination during handling and processing may lead to false positive results.

[Precautions for storage]

(1) Store the product below -20°C.

(2) Do not use products beyond their expiration date.

[Precautions when test]

(1) Before use, visually inspect each component of the product for leaks. Do not use if a leak is detected.

(2) Do not mix this product with products from other lots during use.

(3) Avoid repeated freezing and thawing of the product, as it may affect test results.

(4) This product is optimized for use with the recommended real-time gene amplification device. Use with other equipment may result in incorrect results.

(5) Take care to prevent contamination of product components by nucleic acid extracts, PCR products, or positive controls. We recommend using filter tips to prevent contamination.

(6) In case of spills onto equipment, clean according to the equipment manual and remove contamination from surfaces.

(7) All samples used in testing are potentially infectious. Use appropriate protective equipment and conduct testing in facilities such as biosafety cabinets of Biosafety Level 2 or higher.

(8) Improper collection, transport, or handling of samples may affect test results. Handle samples with caution.

[Precautions when determining results]

(1) Do not utilize test results obtained with this product for disease diagnosis. Diagnosis should be based on test results, clinical findings, and other clinical data evaluated by a healthcare professional.

(2) A negative result alone may not exclude Dengue Virus infection if the viral concentration in the sample is below the detection limit or if amplification inhibitors are present.

(3) Sensitivity may decrease if there are mutations in the primer

and probe binding regions included in the product.

(4) This test does not exclude diseases caused by other bacterial or viral pathogens.

[Precautions for disposal]

(1) Dispose of products and samples following guidelines and methods outlined in the Waste Management Act.

9. Creation and revision date

First creation date : 2024.08.01

[Product name] High-risk infectious agent genetic testing reagent

[Classification number (grade)] K05030.01(3)

[Model name] Asan qPCR Test® Dengue Virus